

#333, 11979 - 40 Street SE , Calgary, AB T2Z 4M3

Customer Service: 877-766-5412

www.optimumenergy.com

BILL TO:

James Galm James Galm 3559 Townley Road

Shaker Heights, OH 44122-

SHIP TO:

James Galm

James Galm

3559 Townley Road

Shaker Heights, OH 44122 5156

	0							
ORDER NO. ORDER DATE		CUSTOMER NO.	PHONE NO.	SHIP VIA	NO. PKGS.			
48090	1/14/2011	g "	2162958458	U3PP	(20) (a) (b) (c) (c) (c) (c) (c) (c) (c) (c) (c) (c			

MESSAGE: Thank you for your business!

Please make cheques payable to Optimum Energy Products Ltd.

Requests for refunds or exchanges will be considered up to (14) fourteen days from date of purchase. Any products to be returned must be unopened and in resalable condition. Software is non-returnable.

Some returns may be subject to a restocking fee.

A return materials authorization number (RMA) must be obtained prior to returning any product to us.

FAX: 403-256-3431 GST: 891086548

	QUANTITY		ITEM NO.	DESCRIPTIO	N COLOR	SIZE	PRICE EA.	TOTAL PRIC
QTY	B/O	SHIPPED						
1	0	1	C3043-CL1.0	Conzerv ELF3234 Panel Mete & Description	er - power	10 14 2		
	0		00040 OL1.0	& energy CZ ELF3234-C	L1.0	14. 4.1	1242	
							S. 128	
						488		
	, •							
	1						,	
uju alimoni i								
						T		
	j have							
,	v.						1	

Thank you for your order!

UFSI-001 (7/05)





## **TEST AND CALIBRATION CERTIFICATE**

Model Number	ELF 3234			
Meter Serial Number	209148 / 46809 - 5010			
Parameter	W / Wh and PF			
Accuracy Class	Class 1.0			

	The transfer of the second
Calibrated On	14-Dec-10

Load	240v, 5A			240v, 1A			240v, 0.2A		
	UPF	0.5 Lag	0.8 Lead	UPF	0.5 Lag	0.8 Lead	UPF	0.5 Lag	0.8 Lead
Permissible Error (%)	± 1.00	± 1.00	± 1.00	± 1.00	± 1.00	± 1.00	± 1.00	± 1.20	± 1.20
Observed Error (%) - W/Wh	0.07	-0.07	0.06	0.00	0.06	-0.06	-0.11	0.35	0.06
Observed Error (%) - PF	-0.08	-0.21	0.08	0.18	-0.07	0.02	0.14	0.18	0.29

Verifi	ed by
Name	SURESH
Signature	SURESH

## Notes

- 1. Reference meter used for calibration has valid calibration certificate traceable to International Standards.
- 2. Schneider Electric Conzerv India Pvt Ltd. recommends calibration interval of 12 months from the date of calibration.
- 3. The Certificate shall not be reproduced except in full, without prior written approval of the issuing authority.
- 4. Accuracy: Applicable to the scope of supply.
  - a) W/Wh: Class 1.0 as per IEC 62052-11 and 62053-21. Class 0.5 (optional) as per IEC 62052-11 and 62053-22.
  - b) PF permissible error limit is same as W permissible error limit in percentage.
  - c) Refer product brochure or leaflet for all the remaining parameters error limits.

Schneider Electric Conzerv India Pvt Ltd www.schneider-electric.co.in

ISO 14001:2004 Certified ISO 9001:2008 Certified

Doc No: TCE 01 - 1